

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
 Data File : VX042149.D
 Acq On : 02 Jul 2024 18:52
 Operator : JC/MD
 Sample : P3119-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW15-20240627

Quant Time: Jul 03 03:38:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	138210	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	213785	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208302	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96310	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70184	44.037	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.080%
35) Dibromofluoromethane	5.385	113	68747	45.320	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.640%
50) Toluene-d8	8.647	98	247861	46.512	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.020%
62) 4-Bromofluorobenzene	11.079	95	90232	45.934	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.860%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1734	2.478	ug/l #	89
30) Chloroform	5.093	83	19373	8.309	ug/l	99
64) Tetrachloroethene	9.275	164	1600	1.055	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
Data File : VX042149.D
Acq On : 02 Jul 2024 18:52
Operator : JC/MD
Sample : P3119-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW15-20240627

Quant Time: Jul 03 03:38:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

