

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
 Data File : VX042128.D
 Acq On : 02 Jul 2024 10:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 03 03:28:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.381	0.469	-23.1	112	0.00
3 P	Chloromethane	0.478	0.472	1.3	91	0.00
4 C	Vinyl Chloride	0.479	0.479	0.0	92	0.00
5 T	Bromomethane	0.237	0.286	-20.7	109	0.00
6 T	Chloroethane	0.241	0.225	6.6	97	0.00
7 T	Trichlorofluoromethane	0.628	0.681	-8.4	102	0.00
8 T	Diethyl Ether	0.295	0.296	-0.3	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.432	0.497	-15.0	104	0.00
10 T	Methyl Iodide	0.537	0.586	-9.1	99	0.00
11 T	Tert butyl alcohol	0.109	0.108	0.9	87	0.00
12 CM	1,1-Dichloroethene	0.431	0.473	-9.7#	99	0.00
13 T	Acrolein	0.123	0.159	-29.3#	116	0.00
14 T	Allyl chloride	0.684	0.735	-7.5	95	0.00
15 T	Acrylonitrile	0.286	0.283	1.0	86	0.00
16 T	Acetone	0.253	0.262	-3.6	101	0.00
17 T	Carbon Disulfide	0.879	1.048	-19.2	114	0.00
18 T	Methyl Acetate	0.691	0.600	13.2	73	0.00
19 T	Methyl tert-butyl Ether	1.425	1.572	-10.3	96	0.00
20 T	Methylene Chloride	0.554	0.523	5.6	93	0.00
21 T	trans-1,2-Dichloroethene	0.438	0.488	-11.4	101	0.00
22 T	Diisopropyl ether	1.446	1.509	-4.4	91	0.00
23 T	Vinyl Acetate	1.235	1.383	-12.0	95	0.00
24 P	1,1-Dichloroethane	0.810	0.863	-6.5	94	0.00
25 T	2-Butanone	0.394	0.390	1.0	86	-0.01
26 T	2,2-Dichloropropane	0.586	0.704	-20.1	107	0.00
27 T	cis-1,2-Dichloroethene	0.548	0.598	-9.1	96	0.00
28 T	Bromochloromethane	0.347	0.310	10.7	83	0.00
29 T	Tetrahydrofuran	0.254	0.249	2.0	84	-0.01
30 C	Chloroform	0.843	0.895	-6.2#	94	-0.01
31 T	Cyclohexane	0.674	0.749	-11.1	102	0.00
32 T	1,1,1-Trichloroethane	0.705	0.778	-10.4	97	0.00
33 S	1,2-Dichloroethane-d4	0.577	0.498	13.7	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	84	0.00
35 S	Dibromofluoromethane	0.355	0.336	5.4	83	0.00
36 T	1,1-Dichloropropene	0.377	0.436	-15.6	101	0.00
37 T	Ethyl Acetate	0.491	0.525	-6.9	87	-0.01
38 T	Carbon Tetrachloride	0.416	0.496	-19.2	101	-0.01
39 T	Methylcyclohexane	0.476	0.593	-24.6	108	0.00
40 TM	Benzene	1.271	1.421	-11.8	96	0.00
41 T	Methacrylonitrile	0.266	0.290	-9.0	88	0.00
42 TM	1,2-Dichloroethane	0.409	0.457	-11.7	94	0.00
43 T	Isopropyl Acetate	0.753	0.802	-6.5	88	0.00
44 TM	Trichloroethene	0.341	0.395	-15.8	98	0.00
45 C	1,2-Dichloropropane	0.312	0.348	-11.5#	93	0.00
46 T	Dibromomethane	0.228	0.258	-13.2	94	0.00
47 T	Bromodichloromethane	0.418	0.476	-13.9	93	0.00
48 T	Methyl methacrylate	0.360	0.394	-9.4	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
 Data File : VX042128.D
 Acq On : 02 Jul 2024 10:39
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 03 03:28:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	87	0.00
50 S	Toluene-d8	1.246	1.126	9.6	79	0.00
51 T	4-Methyl-2-Pentanone	0.519	0.531	-2.3	82	0.00
52 CM	Toluene	0.809	0.913	-12.9#	94	0.00
53 T	t-1,3-Dichloropropene	0.409	0.494	-20.8	97	0.00
54 T	cis-1,3-Dichloropropene	0.458	0.549	-19.9	96	0.00
55 T	1,1,2-Trichloroethane	0.337	0.371	-10.1	90	0.00
56 T	Ethyl methacrylate	0.524	0.599	-14.3	89	0.00
57 T	1,3-Dichloropropane	0.545	0.598	-9.7	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.244	0.265	-8.6	84	0.00
59 T	2-Hexanone	0.397	0.411	-3.5	82	0.00
60 T	Dibromochloromethane	0.366	0.427	-16.7	92	0.00
61 T	1,2-Dibromoethane	0.350	0.395	-12.9	93	0.00
62 S	4-Bromofluorobenzene	0.459	0.441	3.9	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	81	0.00
64 T	Tetrachloroethene	0.364	0.429	-17.9	101	0.00
65 PM	Chlorobenzene	1.040	1.172	-12.7	93	0.00
66 T	1,1,1,2-Tetrachloroethane	0.364	0.427	-17.3	92	0.00
67 C	Ethyl Benzene	1.630	1.905	-16.9#	94	0.00
68 T	m/p-Xylenes	0.658	0.770	-17.0	95	0.00
69 T	o-Xylene	0.651	0.752	-15.5	92	0.00
70 T	Styrene	1.092	1.300	-19.0	92	0.00
71 P	Bromoform	0.326	0.392	-20.2	91	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
73 T	Isopropylbenzene	3.040	3.441	-13.2	94	0.00
74 T	N-amyl acetate	1.350	1.422	-5.3	86	0.00
75 P	1,1,2,2-Tetrachloroethane	1.130	1.146	-1.4	87	0.00
76 T	1,2,3-Trichloropropane	0.908	0.931	-2.5	89	0.00
77 T	Bromobenzene	0.905	0.989	-9.3	92	0.00
78 T	n-propylbenzene	3.299	3.796	-15.1	94	0.00
79 T	2-Chlorotoluene	2.125	2.303	-8.4	92	0.00
80 T	1,3,5-Trimethylbenzene	2.532	2.878	-13.7	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.320	0.358	-11.9	91	0.00
82 T	4-Chlorotoluene	2.349	2.635	-12.2	92	0.00
83 T	tert-Butylbenzene	2.779	3.051	-9.8	91	0.00
84 T	1,2,4-Trimethylbenzene	2.541	2.919	-14.9	93	0.00
85 T	sec-Butylbenzene	3.143	3.630	-15.5	94	0.00
86 T	p-Isopropyltoluene	2.745	3.222	-17.4	95	0.00
87 T	1,3-Dichlorobenzene	1.592	1.774	-11.4	93	0.00
88 T	1,4-Dichlorobenzene	1.633	1.779	-8.9	93	0.00
89 T	n-Butylbenzene	2.113	2.522	-19.4	96	0.00
90 T	Hexachloroethane	0.467	0.534	-14.3	94	0.00
91 T	1,2-Dichlorobenzene	1.629	1.805	-10.8	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.224	0.241	-7.6	86	0.00
93 T	1,2,4-Trichlorobenzene	1.067	1.258	-17.9	96	0.00
94 T	Hexachlorobutadiene	0.539	0.623	-15.6	99	0.00
95 T	Naphthalene	3.289	3.788	-15.2	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070224\
Data File : VX042128.D
Acq On : 02 Jul 2024 10:39
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 03 03:28:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.125	1.292	-14.8	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 5