

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010635.D
 Acq On : 03 Jul 2019 13:18
 Operator : JC/SP
 Sample : K3662-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-TB1-20190702

Quant Time: Jul 04 04:01:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	243279	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	377046	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	337433	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	153564	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	119975	52.60	ug/l	0.00
Spiked Amount			Recovery	=	105.20%	
35) Dibromofluoromethane	5.50	113	114329	49.56	ug/l	0.00
Spiked Amount			Recovery	=	99.12%	
50) Toluene-d8	8.71	98	440666	49.93	ug/l	0.00
Spiked Amount			Recovery	=	99.86%	
62) 4-Bromofluorobenzene	11.13	95	147508	46.29	ug/l	0.00
Spiked Amount			Recovery	=	92.58%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070319\
Data File : VX010635.D
Acq On : 03 Jul 2019 13:18
Operator : JC/SP
Sample : K3662-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-TB1-20190702

Quant Time: Jul 04 04:01:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

