

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010636.D
 Acq On : 03 Jul 2019 13:41
 Operator : JC/SP
 Sample : K3667-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-190702

Quant Time: Jul 04 04:03:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	247713	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	383175	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	347607	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	158691	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	123824	53.31	ug/l	0.00
Spiked Amount						
			Recovery	=	106.62%	
35) Dibromofluoromethane	5.50	113	116777	49.81	ug/l	0.00
Spiked Amount						
			Recovery	=	99.62%	
50) Toluene-d8	8.71	98	452777	50.48	ug/l	0.00
Spiked Amount						
			Recovery	=	100.96%	
62) 4-Bromofluorobenzene	11.13	95	152882	47.21	ug/l	0.00
Spiked Amount						
			Recovery	=	94.42%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070319\
Data File : VX010636.D
Acq On : 03 Jul 2019 13:41
Operator : JC/SP
Sample : K3667-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-190702

Quant Time: Jul 04 04:03:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

