

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010638.D
 Acq On : 03 Jul 2019 14:28
 Operator : JC/SP
 Sample : K3635-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Jul 08 14:44:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	244553	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	373695	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	340230	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	155329	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	121694	53.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.14%	
35) Dibromofluoromethane	5.50	113	117009	51.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.36%	
50) Toluene-d8	8.71	98	444442	50.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.62%	
62) 4-Bromofluorobenzene	11.13	95	149331	47.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.58%	
Target Compounds						
16) Acetone	2.45	43	3740	3.255	ug/l	94
17) Carbon Disulfide	2.57	76	2073	0.365	ug/l #	85
40) Benzene	6.15	78	2568	0.278	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070319\
Data File : VX010638.D
Acq On : 03 Jul 2019 14:28
Operator : JC/SP
Sample : K3635-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Jul 08 14:44:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

