

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010643.D
 Acq On : 03 Jul 2019 16:25
 Operator : JC/SP
 Sample : K3635-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-B

Quant Time: Jul 04 04:25:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	243722	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	374827	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	341478	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	159255	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	120106	52.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.12%	
35) Dibromofluoromethane	5.50	113	114500	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
50) Toluene-d8	8.71	98	443570	50.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.12%	
62) 4-Bromofluorobenzene	11.13	95	149074	47.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.12%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070319\
Data File : VX010643.D
Acq On : 03 Jul 2019 16:25
Operator : JC/SP
Sample : K3635-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-B

Quant Time: Jul 04 04:25:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

