

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010647.D
 Acq On : 03 Jul 2019 17:58
 Operator : JC/SP
 Sample : K3667-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 OUTFALL001-190702

Quant Time: Jul 04 04:39:00 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	231973	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	358337	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	324163	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	146558	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	116157	53.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.82%	
35) Dibromofluoromethane	5.50	113	110054	50.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.40%	
50) Toluene-d8	8.71	98	424532	50.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.22%	
62) 4-Bromofluorobenzene	11.13	95	142089	46.92	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.84%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	18732	7.496	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070319\
Data File : VX010647.D
Acq On : 03 Jul 2019 17:58
Operator : JC/SP
Sample : K3667-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
OUTFALL001-190702

Quant Time: Jul 04 04:39:00 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

