

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010648.D
 Acq On : 03 Jul 2019 18:21
 Operator : JC/SP
 Sample : K3636-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SP19-JMW0991

Quant Time: Jul 04 04:55:29 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	229991	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	357615	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	327505	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	148727	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	115080	53.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.74%	
35) Dibromofluoromethane	5.50	113	109175	49.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.80%	
50) Toluene-d8	8.71	98	425460	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
62) 4-Bromofluorobenzene	11.13	95	141501	46.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.64%	
Target Compounds						
8) Diethyl Ether	2.19	74	5736	4.948	ug/l	95
27) cis-1,2-Dichloroethene	4.59	96	1052	0.425	ug/l	87
40) Benzene	6.14	78	3301	0.373	ug/l #	82
44) Trichloroethene	7.21	130	1183	0.447	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070319\
Data File : VX010648.D
Acq On : 03 Jul 2019 18:21
Operator : JC/SP
Sample : K3636-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SP19-JMW0991

Quant Time: Jul 04 04:55:29 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

