

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010654.D
 Acq On : 03 Jul 2019 20:42
 Operator : JC/SP
 Sample : K3666-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MWHR3-14-070219

Quant Time: Jul 04 05:19:22 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	226717	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	353436	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	321900	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	147959	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	114384	53.81	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.62%	
35) Dibromofluoromethane	5.50	113	108117	50.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.00%	
50) Toluene-d8	8.71	98	418908	50.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.26%	
62) 4-Bromofluorobenzene	11.13	95	140881	47.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.34%	
Target Compounds						
16) Acetone	2.44	43	7020	6.589	ug/l #	87
65) Chlorobenzene	10.13	112	16777	2.567	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070319\
Data File : VX010654.D
Acq On : 03 Jul 2019 20:42
Operator : JC/SP
Sample : K3666-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MWHR3-14-070219

Quant Time: Jul 04 05:19:22 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

