

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010659.D
 Acq On : 03 Jul 2019 22:38
 Operator : JC/SP
 Sample : K3618-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HD-01-070219

Quant Time: Jul 04 05:37:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	240658	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	379836	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	351580	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	166796	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	119543	52.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.96%	
35) Dibromofluoromethane	5.50	113	115339	49.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.26%	
50) Toluene-d8	8.71	98	459781	51.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.42%	
62) 4-Bromofluorobenzene	11.13	95	154627	48.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.34%	
Target Compounds						
16) Acetone	2.44	43	35278	31.196	ug/l	99
20) Methylene Chloride	2.86	84	6101	2.622	ug/l #	95
25) 2-Butanone	4.68	43	9544	5.754	ug/l	89
29) Tetrahydrofuran	5.13	42	38641	37.608	ug/l	99
43) Isopropyl Acetate	6.45	43	14588	2.840	ug/l	99
95) Naphthalene	13.83	128	94178	8.068	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070319\
Data File : VX010659.D
Acq On : 03 Jul 2019 22:38
Operator : JC/SP
Sample : K3618-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HD-01-070219

Quant Time: Jul 04 05:37:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

