

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010673.D
 Acq On : 04 Jul 2019 07:28
 Operator : JC/SP
 Sample : K3669-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW218-190702

Quant Time: Jul 05 02:14:56 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	239931	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	373725	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	340730	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	155800	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	120080	53.38	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.76%	
35) Dibromofluoromethane	5.50	113	113856	49.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.58%	
50) Toluene-d8	8.71	98	443573	50.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.42%	
62) 4-Bromofluorobenzene	11.13	95	147701	46.77	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.54%	
Target Compounds						
4) Vinyl Chloride	1.41	62	8251	3.878	ug/l	99
16) Acetone	2.45	43	5311	4.711	ug/l	94
27) cis-1,2-Dichloroethene	4.60	96	8634	3.340	ug/l	96
40) Benzene	6.15	78	9895	1.069	ug/l	93
65) Chlorobenzene	10.14	112	62115	8.979	ug/l	99
83) tert-Butylbenzene	11.77	119	3781	0.433	ug/l	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070319\
Data File : VX010673.D
Acq On : 04 Jul 2019 07:28
Operator : JC/SP
Sample : K3669-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW218-190702

Quant Time: Jul 05 02:14:56 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

