

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010660.D
 Acq On : 03 Jul 2019 23:02
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 04 05:38:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	73	0.00
2 T	Dichlorodifluoromethane	0.348	0.438	-25.9#	95	0.00
3 P	Chloromethane	0.405	0.438	-8.1	82	0.00
4 C	Vinyl Chloride	0.443	0.498	-12.4#	87	0.00
5 T	Bromomethane	0.216	0.250	-15.7	90	0.00
6 T	Chloroethane	0.262	0.306	-16.8	88	0.00
7 T	Trichlorofluoromethane	0.703	0.844	-20.1#	89	0.00
8 T	Diethyl Ether	0.252	0.292	-15.9	86	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.426	0.474	-11.3	85	0.00
10 T	Methyl Iodide	0.608	0.608	0.0	70	0.00
11 T	Tert butyl alcohol	0.123	0.104	15.4	65	0.00
12 CM	1,1-Dichloroethene	0.433	0.478	-10.4#	82	0.00
13 T	Acrolein	0.081	0.051	37.0#	49#	0.00
14 T	Allyl chloride	0.615	0.678	-10.2	80	0.00
15 T	Acrylonitrile	0.243	0.266	-9.5	80	0.00
16 T	Acetone	0.235	0.211	10.2	63	0.00
17 T	Carbon Disulfide	1.163	1.209	-4.0	78	0.00
18 T	Methyl Acetate	0.468	0.679	-45.1#	108	0.00
19 T	Methyl tert-butyl Ether	1.363	1.543	-13.2	84	0.00
20 T	Methylene Chloride	0.484	0.533	-10.1	83	0.00
21 T	trans-1,2-Dichloroethene	0.466	0.508	-9.0	82	0.00
22 T	Diisopropyl ether	1.286	1.451	-12.8	84	0.00
23 T	Vinyl Acetate	1.111	1.066	4.1	68	0.00
24 P	1,1-Dichloroethane	0.744	0.830	-11.6	82	0.00
25 T	2-Butanone	0.345	0.355	-2.9	74	0.00
26 T	2,2-Dichloropropane	0.646	0.536	17.0	62	0.00
27 T	cis-1,2-Dichloroethene	0.539	0.588	-9.1	81	0.00
28 T	Bromochloromethane	0.356	0.362	-1.7	77	0.00
29 T	Tetrahydrofuran	0.213	0.234	-9.9	80	0.00
30 C	Chloroform	0.800	0.895	-11.9#	84	0.00
31 T	Cyclohexane	0.677	0.764	-12.9	83	0.00
32 T	1,1,1-Trichloroethane	0.706	0.789	-11.8	82	0.00
33 S	1,2-Dichloroethane-d4	0.469	0.488	-4.1	76	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	75	0.00
35 S	Dibromofluoromethane	0.306	0.312	-2.0	75	0.00
36 T	1,1-Dichloropropene	0.397	0.433	-9.1	84	0.00
37 T	Ethyl Acetate	0.409	0.438	-7.1	78	0.00
38 T	Carbon Tetrachloride	0.430	0.458	-6.5	79	0.00
39 T	Methylcyclohexane	0.522	0.550	-5.4	80	0.00
40 TM	Benzene	1.238	1.348	-8.9	82	0.00
41 T	Methacrylonitrile	0.224	0.251	-12.1	84	0.00
42 TM	1,2-Dichloroethane	0.377	0.443	-17.5	86	0.00
43 T	Isopropyl Acetate	0.676	0.721	-6.7	79	0.00
44 TM	Trichloroethene	0.370	0.383	-3.5	78	0.00
45 C	1,2-Dichloropropane	0.313	0.338	-8.0#	82	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
 Data File : VX010660.D
 Acq On : 03 Jul 2019 23:02
 Operator : JC/SP
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 04 05:38:41 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 20 03:31:56 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.224	0.247	-10.3	83	0.00
47 T	Bromodichloromethane	0.411	0.443	-7.8	79	0.00
48 T	Methyl methacrylate	0.314	0.359	-14.3	83	0.00
49 T	1,4-Dioxane	0.009	0.008	11.1	68	0.00
50 S	Toluene-d8	1.170	1.126	3.8	70	0.00
51 T	4-Methyl-2-Pentanone	0.433	0.470	-8.5	80	0.00
52 CM	Toluene	0.815	0.878	-7.7#	81	0.00
53 T	t-1,3-Dichloropropene	0.461	0.468	-1.5	75	0.00
54 T	cis-1,3-Dichloropropene	0.501	0.528	-5.4	77	0.00
55 T	1,1,2-Trichloroethane	0.336	0.365	-8.6	81	0.00
56 T	Ethyl methacrylate	0.509	0.548	-7.7	78	0.00
57 T	1,3-Dichloropropane	0.522	0.568	-8.8	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.242	0.233	3.7	69	0.00
59 T	2-Hexanone	0.340	0.351	-3.2	75	0.00
60 T	Dibromochloromethane	0.376	0.392	-4.3	76	0.00
61 T	1,2-Dibromoethane	0.363	0.397	-9.4	81	0.00
62 S	4-Bromofluorobenzene	0.423	0.406	4.0	70	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	73	0.00
64 T	Tetrachloroethene	0.404	0.386	4.5	73	0.00
65 PM	Chlorobenzene	1.015	1.078	-6.2	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.370	0.397	-7.3	77	0.00
67 C	Ethyl Benzene	1.714	1.856	-8.3#	80	0.00
68 T	m/p-Xylenes	0.664	0.713	-7.4	79	0.00
69 T	o-Xylene	0.653	0.695	-6.4	79	0.00
70 T	Styrene	1.095	1.186	-8.3	78	0.00
71 P	Bromoform	0.322	0.327	-1.6	70	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
73 T	Isopropylbenzene	3.389	3.647	-7.6	79	0.00
74 T	N-amyl acetate	1.281	1.121	12.5	61	0.00
75 P	1,1,2,2-Tetrachloroethane	1.122	1.207	-7.6	78	0.00
76 T	1,2,3-Trichloropropane	0.911	0.985	-8.1	78	0.00
77 T	Bromobenzene	0.946	0.986	-4.2	77	0.00
78 T	n-propylbenzene	3.725	4.065	-9.1	78	0.00
79 T	2-Chlorotoluene	2.218	2.375	-7.1	78	0.00
80 T	1,3,5-Trimethylbenzene	2.794	2.990	-7.0	77	0.00
81 T	trans-1,4-Dichloro-2-butene	0.391	0.346	11.5	62	0.00
82 T	4-Chlorotoluene	2.524	2.765	-9.5	78	0.00
83 T	tert-Butylbenzene	2.804	3.032	-8.1	78	0.00
84 T	1,2,4-Trimethylbenzene	2.807	3.052	-8.7	77	0.00
85 T	sec-Butylbenzene	3.293	3.514	-6.7	77	0.00
86 T	p-Isopropyltoluene	3.046	3.249	-6.7	76	0.00
87 T	1,3-Dichlorobenzene	1.652	1.715	-3.8	77	0.00
88 T	1,4-Dichlorobenzene	1.682	1.704	-1.3	76	0.00
89 T	n-Butylbenzene	2.588	2.732	-5.6	74	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070319\
Data File : VX010660.D
Acq On : 03 Jul 2019 23:02
Operator : JC/SP
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 04 05:38:41 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061919W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 20 03:31:56 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.527	0.535	-1.5	70	0.00
91 T	1,2-Dichlorobenzene	1.642	1.697	-3.3	76	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.242	0.244	-0.8	75	0.00
93 T	1,2,4-Trichlorobenzene	1.139	1.159	-1.8	74	0.00
94 T	Hexachlorobutadiene	0.556	0.536	3.6	70	0.00
95 T	Naphthalene	3.499	3.645	-4.2	74	0.00
96 T	1,2,3-Trichlorobenzene	1.132	1.160	-2.5	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6