

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070324\
 Data File : VX042156.D
 Acq On : 03 Jul 2024 12:02
 Operator : JC/MD
 Sample : VX0703MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0703MBL01

Quant Time: Jul 04 07:21:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	140848	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	216863	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	213316	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105345	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68910	42.428	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.860%
35) Dibromofluoromethane	5.385	113	69435	45.124	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.240%
50) Toluene-d8	8.647	98	254431	47.067	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.140%
62) 4-Bromofluorobenzene	11.079	95	95651	48.001	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070324\
Data File : VX042156.D
Acq On : 03 Jul 2024 12:02
Operator : JC/MD
Sample : VX0703MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0703MBL01

Quant Time: Jul 04 07:21:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

