

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070324\
 Data File : VX042157.D
 Acq On : 03 Jul 2024 12:26
 Operator : JC/MD
 Sample : VX0703WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0703WBL01

Quant Time: Jul 04 07:22:02 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	145272	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	226128	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225337	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108493	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	73819	44.067	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.140%
35) Dibromofluoromethane	5.385	113	70990	44.244	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.480%
50) Toluene-d8	8.647	98	264794	46.977	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.960%
62) 4-Bromofluorobenzene	11.079	95	98982	47.637	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.280%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070324\
Data File : VX042157.D
Acq On : 03 Jul 2024 12:26
Operator : JC/MD
Sample : VX0703WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0703WBL01

Quant Time: Jul 04 07:22:02 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

