

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070324\
 Data File : VX042169.D
 Acq On : 03 Jul 2024 17:09
 Operator : JC/MD
 Sample : P3163-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOWER-1

Quant Time: Jul 04 07:29:08 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143321	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	219795	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216580	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102830	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	70445	42.625	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.240%
35) Dibromofluoromethane	5.385	113	69066	44.286	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.580%
50) Toluene-d8	8.647	98	255717	46.674	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.340%
62) 4-Bromofluorobenzene	11.079	95	96095	47.581	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.160%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	10860	14.965	ug/l	91
20) Methylene Chloride	2.782	84	1229	0.773	ug/l #	84
30) Chloroform	5.093	83	85947	35.549	ug/l	95
47) Bromodichloromethane	7.824	83	19867	10.814	ug/l #	96
53) t-1,3-Dichloropropene	9.000	75	1103	1.463	ug/l	95
60) Dibromochloromethane	9.525	129	7580	4.710	ug/l	99
71) Bromoform	10.805	173	770	1.394	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070324\
Data File : VX042169.D
Acq On : 03 Jul 2024 17:09
Operator : JC/MD
Sample : P3163-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOWER-1

Quant Time: Jul 04 07:29:08 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

