

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070325\
 Data File : VX046870.D
 Acq On : 03 Jul 2025 08:42
 Operator : JC/MD
 Sample : VX0703MBL01
 Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0703MBL01

Quant Time: Jul 04 01:25:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	347322	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	593862	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	536305	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	269022	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	225335	49.109	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.220%
35) Dibromofluoromethane	5.391	113	192478	47.748	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.500%
50) Toluene-d8	8.647	98	711018	49.992	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.980%
62) 4-Bromofluorobenzene	11.079	95	266944	49.384	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070325\
Data File : VX046870.D
Acq On : 03 Jul 2025 08:42
Operator : JC/MD
Sample : VX0703MBL01
Misc : 5.00g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0703MBL01

Quant Time: Jul 04 01:25:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

