

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070325\
 Data File : VX046872.D
 Acq On : 03 Jul 2025 09:35
 Operator : JC/MD
 Sample : Q2458-10
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB-06272025

Quant Time: Jul 04 01:26:34 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	397171	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	689071	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	628783	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	317895	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	265228	50.548	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	101.100%
35) Dibromofluoromethane	5.391	113	223276	47.735	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.460%
50) Toluene-d8	8.647	98	821236	49.763	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.520%
62) 4-Bromofluorobenzene	11.079	95	313594	49.998	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070325\
Data File : VX046872.D
Acq On : 03 Jul 2025 09:35
Operator : JC/MD
Sample : Q2458-10
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB-06272025

Quant Time: Jul 04 01:26:34 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

