

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070325\
 Data File : VX046878.D
 Acq On : 03 Jul 2025 11:46
 Operator : JC/MD
 Sample : Q2455-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FIELD-BLANK

Quant Time: Jul 04 01:29:46 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	419216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	727439	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	669234	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	336823	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	278540	50.294	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.580%
35) Dibromofluoromethane	5.397	113	236123	47.819	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.640%
50) Toluene-d8	8.647	98	865532	49.681	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.360%
62) 4-Bromofluorobenzene	11.079	95	330095	49.853	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070325\
Data File : VX046878.D
Acq On : 03 Jul 2025 11:46
Operator : JC/MD
Sample : Q2455-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FIELD-BLANK

Quant Time: Jul 04 01:29:46 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

