

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070325\
 Data File : VX046879.D
 Acq On : 03 Jul 2025 12:07
 Operator : JC/MD
 Sample : Q2447-07ME
 Misc : 3.00g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LAW-25-0100ME

Quant Time: Jul 04 01:30:03 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	396544	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	672060	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	611341	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	308520	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	253531	48.395	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	96.800%		
35) Dibromofluoromethane	5.391	113	209749	45.978	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	91.960%		
50) Toluene-d8	8.647	98	800768	49.751	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.500%		
62) 4-Bromofluorobenzene	11.079	95	301787	49.334	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	98.660%		
Target Compounds						
16) Acetone	2.398	43	7494	4.605	ug/l	96
39) Methylcyclohexane	7.385	83	10136	1.316	ug/l #	92
40) Benzene	6.044	78	6132	0.329	ug/l	95
52) Toluene	8.720	92	415725	36.036	ug/l	98
67) Ethyl Benzene	10.195	91	19281	0.852	ug/l	96
68) m/p-Xylenes	10.299	106	18589	2.179	ug/l	96
69) o-Xylene	10.640	106	5052	0.613	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070325\
Data File : VX046879.D
Acq On : 03 Jul 2025 12:07
Operator : JC/MD
Sample : Q2447-07ME
Misc : 3.00g/10.0mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LAW-25-0100ME

Quant Time: Jul 04 01:30:03 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

