

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070325\
 Data File : VX046886.D
 Acq On : 03 Jul 2025 14:37
 Operator : JC/MD
 Sample : Q2503-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GCAP2W

Quant Time: Jul 04 01:32:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 03 05:51:11 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	529972	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	893199	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	813566	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	408234	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	335135	47.867	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.740%
35) Dibromofluoromethane	5.397	113	285078	47.019	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.040%
50) Toluene-d8	8.647	98	1067887	49.920	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.840%
62) 4-Bromofluorobenzene	11.079	95	407032	50.065	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.120%
Target Compounds						
					Qvalue	
3) Chloromethane	1.307	50	13721	2.478	ug/l	96
31) Cyclohexane	5.483	56	18536	1.852	ug/l	99
39) Methylcyclohexane	7.385	83	20561	2.009	ug/l	99
52) Toluene	8.720	92	7353	0.480	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070325\
Data File : VX046886.D
Acq On : 03 Jul 2025 14:37
Operator : JC/MD
Sample : Q2503-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GCAP2W

Quant Time: Jul 04 01:32:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X070225W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 03 05:51:11 2025
Response via : Initial Calibration

