

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070424\
 Data File : VX042196.D
 Acq On : 04 Jul 2024 04:45
 Operator : JC/MD
 Sample : P3088-17ME
 Misc : 4.52g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 A4D12ME

Quant Time: Jul 04 05:19:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML070324WMA.M
 Quant Title : VOC Analysis
 QLast Update : Thu Jul 04 05:08:27 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
 Supervised By :Semsettin Yesilyurt 07/05/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	222057	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	203323	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	100203	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.404	65	40522	31.351	ug/L	0.04
Spiked Amount 50.000	Range 60 - 135		Recovery =	62.700%		
7) Chloroethane-d5	1.746	69	24449m	30.043	ug/L	0.10
Spiked Amount 50.000	Range 70 - 130		Recovery =	60.080%#		
11) 1,1-Dichloroethene-d2	2.270	65	16238	28.705	ug/L	-0.02
Spiked Amount 50.000	Range 60 - 125		Recovery =	57.420%#		
21) 2-Butanone-d5	4.495	46	52082	55.916	ug/L	0.03
Spiked Amount 100.000	Range 40 - 130		Recovery =	55.920%		
24) Chloroform-d	5.056	84	77717	29.905	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	59.820%#		
26) 1,2-Dichloroethane-d4	5.952	65	51273	33.426	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	66.860%#		
32) Benzene-d6	5.964	84	173594	33.031	ug/L	0.00
Spiked Amount 50.000	Range 70 - 125		Recovery =	66.060%#		
36) 1,2-Dichloropropane-d6	7.306	67	52225	33.021	ug/L	0.00
Spiked Amount 50.000	Range 70 - 120		Recovery =	66.040%#		
41) Toluene-d8	8.647	98	155935	31.840	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	63.680%#		
43) trans-1,3-Dichloroprop...	8.952	79	17505	25.609	ug/L	0.00
Spiked Amount 50.000	Range 60 - 125		Recovery =	51.220%#		
47) 2-Hexanone-d5	9.391	63	42687	59.460	ug/L	0.00
Spiked Amount 100.000	Range 45 - 130		Recovery =	59.460%		
56) 1,1,2,2-Tetrachloroeth...	11.195	84	68567	29.325	ug/L	0.00
Spiked Amount 50.000	Range 65 - 120		Recovery =	58.640%#		
66) 1,2-Dichlorobenzene-d4	12.323	152	65548	33.679	ug/L	0.00
Spiked Amount 50.000	Range 80 - 120		Recovery =	67.360%#		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070424\
Data File : VX042196.D
Acq On : 04 Jul 2024 04:45
Operator : JC/MD
Sample : P3088-17ME
Misc : 4.52g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
A4D12ME

Quant Time: Jul 04 05:19:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXLM070324WMA.M
Quant Title : VOC Analysis
QLast Update : Thu Jul 04 05:08:27 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Mahesh Dadoda 07/05/2024
Supervised By :Semsettin Yesilyurt 07/05/2024

