

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070518\
 Data File : VX003170.D
 Acq On : 05 Jul 2018 14:20
 Operator : JC/MD
 Sample : J3754-08DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S10-0145DL

Quant Time: Jul 05 17:32:51 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	271428	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	376336	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	345887	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	194211	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	172955	46.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.20%	
35) Dibromofluoromethane	5.51	113	144461	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
50) Toluene-d8	8.72	98	519591	47.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.48%	
62) 4-Bromofluorobenzene	11.14	95	180918	43.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.22%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
40) Benzene	6.15	78	695554	58.01	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

