

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070518\
 Data File : VX003194.D
 Acq On : 06 Jul 2018 01:54
 Operator : JC/MD
 Sample : J3844-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 34 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 CPTB01-180702

Quant Time: Jul 06 11:17:50 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	287755	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	405760	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	365725	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	204562	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	183600	46.16	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.32%	
35) Dibromofluoromethane	5.51	113	151701	47.21	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.42%	
50) Toluene-d8	8.72	98	555247	47.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.64%	
62) 4-Bromofluorobenzene	11.14	95	193603	42.78	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.56%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070518\
Data File : VX003194.D
Acq On : 06 Jul 2018 01:54
Operator : JC/MD
Sample : J3844-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 34 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
CPTB01-180702

Quant Time: Jul 06 11:17:50 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

