

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070518\
 Data File : VX003196.D
 Acq On : 06 Jul 2018 02:47
 Operator : JC/MD
 Sample : J3844-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP022MW001-180702

Quant Time: Jul 06 11:19:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	270459	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	374768	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	344738	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.09	152	192241	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	170641	45.65	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.30%	
35) Dibromofluoromethane	5.51	113	145022	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.72%	
50) Toluene-d8	8.72	98	519485	47.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.86%	
62) 4-Bromofluorobenzene	11.14	95	180859	43.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.54%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
42) 1,2-Dichloroethane	6.20	62	17472	3.62	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070518\
Data File : VX003196.D
Acq On : 06 Jul 2018 02:47
Operator : JC/MD
Sample : J3844-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP022MW001-180702

Quant Time: Jul 06 11:19:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

