

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070518\
 Data File : VX003199.D
 Acq On : 06 Jul 2018 04:07
 Operator : JC/MD
 Sample : J3844-07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP022RW358-180702

Quant Time: Jul 06 04:48:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	271444	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	378026	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	349178	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	199077	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	172571	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
35) Dibromofluoromethane	5.51	113	144808	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
50) Toluene-d8	8.72	98	519705	47.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.08%	
62) 4-Bromofluorobenzene	11.14	95	187323	44.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.86%	
Target Compounds						
4) Vinyl Chloride	1.40	62	8555	2.79	ug/l	# 67
65) Chlorobenzene	10.14	112	4768	0.54	ug/l	93
87) 1,3-Dichlorobenzene	12.03	146	1416	0.20	ug/l	96
91) 1,2-Dichlorobenzene	12.40	146	1574	0.22	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070518\
Data File : VX003199.D
Acq On : 06 Jul 2018 04:07
Operator : JC/MD
Sample : J3844-07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP022RW358-180702

Quant Time: Jul 06 04:48:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

