

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
 Data File : VX029915.D
 Acq On : 05 Jul 2022 13:08
 Operator : JC/MD
 Sample : PB145991ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145991ZHE#01

Quant Time: Jul 06 06:54:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	221252	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	378649	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	322292	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	137734	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	120301	42.954	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	85.900%
35) Dibromofluoromethane	5.391	113	57166	24.158	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	48.320%#
50) Toluene-d8	8.653	98	427642	48.172	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.340%
62) 4-Bromofluorobenzene	11.079	95	143321	45.673	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.340%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	49187	35.898	ug/l	100
20) Methylene Chloride	2.794	84	20179	6.973	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
Data File : VX029915.D
Acq On : 05 Jul 2022 13:08
Operator : JC/MD
Sample : PB145991ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145991ZHE#01

Quant Time: Jul 06 06:54:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

