

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
 Data File : VX029916.D
 Acq On : 05 Jul 2022 13:31
 Operator : JC/MD
 Sample : PB145991ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145991ZHE#02

Quant Time: Jul 06 06:54:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	214945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	362583	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	303473	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	121327	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	114186	41.967	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	83.940%
35) Dibromofluoromethane	5.391	113	54770	24.171	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	48.340%#
50) Toluene-d8	8.653	98	411137	48.365	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	96.740%
62) 4-Bromofluorobenzene	11.085	95	129888	43.226	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.460%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	47890	35.977	ug/l	90
20) Methylene Chloride	2.788	84	20244	7.201	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
Data File : VX029916.D
Acq On : 05 Jul 2022 13:31
Operator : JC/MD
Sample : PB145991ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145991ZHE#02

Quant Time: Jul 06 06:54:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

