

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
 Data File : VX029918.D
 Acq On : 05 Jul 2022 14:18
 Operator : JC/MD
 Sample : PB145991ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145991ZHE#04

Quant Time: Jul 06 06:55:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.555	168	207704	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	358250	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	306211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128542	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	113594	43.205	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	86.420%		
35) Dibromofluoromethane	5.391	113	53067	23.703	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	47.400%#		
50) Toluene-d8	8.652	98	408098	48.588	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.180%		
62) 4-Bromofluorobenzene	11.085	95	133998	45.133	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	90.260%		
Target Compounds						
16) Acetone	2.385	43	48885	38.005	ug/l	99
20) Methylene Chloride	2.794	84	19927	7.335	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
Data File : VX029918.D
Acq On : 05 Jul 2022 14:18
Operator : JC/MD
Sample : PB145991ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145991ZHE#04

Quant Time: Jul 06 06:55:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

