

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
 Data File : VX029920.D
 Acq On : 05 Jul 2022 15:05
 Operator : JC/MD
 Sample : PB145991ZHE#06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145991ZHE#06

Quant Time: Jul 06 06:55:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	210318	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	362541	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	314739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	134769	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	116307	43.687	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.380%
35) Dibromofluoromethane	5.391	113	104591	46.163	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.320%
50) Toluene-d8	8.653	98	413095	48.601	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.200%
62) 4-Bromofluorobenzene	11.085	95	138812	46.201	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.400%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	22991	17.652	ug/l	100
20) Methylene Chloride	2.788	84	11685	4.248	ug/l	90
37) Ethyl Acetate	4.727	43	14627	3.549	ug/l	96
43) Isopropyl Acetate	6.348	43	62329	10.043	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
Data File : VX029920.D
Acq On : 05 Jul 2022 15:05
Operator : JC/MD
Sample : PB145991ZHE#06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145991ZHE#06

Quant Time: Jul 06 06:55:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

