

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
 Data File : VX029922.D
 Acq On : 05 Jul 2022 15:52
 Operator : JC/MD
 Sample : PB145991ZHE#08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145991ZHE#08

Quant Time: Jul 06 06:56:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	208204	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	363135	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	314288	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	131254	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	115427	43.797	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.600%
35) Dibromofluoromethane	5.391	113	103659	45.677	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.360%
50) Toluene-d8	8.653	98	415076	48.754	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.500%
62) 4-Bromofluorobenzene	11.085	95	136340	45.304	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.600%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	22734	17.632	ug/l	97
20) Methylene Chloride	2.788	84	11182	4.106	ug/l	93
37) Ethyl Acetate	4.733	43	13930	3.374	ug/l #	93
43) Isopropyl Acetate	6.349	43	62483	10.051	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
Data File : VX029922.D
Acq On : 05 Jul 2022 15:52
Operator : JC/MD
Sample : PB145991ZHE#08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145991ZHE#08

Quant Time: Jul 06 06:56:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

