

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
 Data File : VX029926.D
 Acq On : 05 Jul 2022 17:26
 Operator : JC/MD
 Sample : PB145991ZHE#12
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145991ZHE#12

Quant Time: Jul 06 06:57:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	204003	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	356882	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	304718	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	130387	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	112447	43.545	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	87.080%
35) Dibromofluoromethane	5.391	113	101876	45.678	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.360%
50) Toluene-d8	8.653	98	398720	47.654	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.300%
62) 4-Bromofluorobenzene	11.085	95	134799	45.577	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.160%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	20917	16.557	ug/l	100
20) Methylene Chloride	2.794	84	11293	4.233	ug/l	94
37) Ethyl Acetate	4.727	43	17936	4.421	ug/l	95
43) Isopropyl Acetate	6.348	43	75965	12.434	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
Data File : VX029926.D
Acq On : 05 Jul 2022 17:26
Operator : JC/MD
Sample : PB145991ZHE#12
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145991ZHE#12

Quant Time: Jul 06 06:57:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

