

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
 Data File : VX029929.D
 Acq On : 05 Jul 2022 18:36
 Operator : JC/MD
 Sample : PB145991ZHE#15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB145991ZHE#15

Quant Time: Jul 06 06:58:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	203099	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	357137	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	315384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	136370	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	114385	44.492	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.980%
35) Dibromofluoromethane	5.391	113	103149	46.215	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	92.440%
50) Toluene-d8	8.653	98	414402	49.493	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.980%
62) 4-Bromofluorobenzene	11.085	95	138303	46.728	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.460%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	20968	16.671	ug/l	98
20) Methylene Chloride	2.794	84	17853	6.721	ug/l	94
37) Ethyl Acetate	4.727	43	38910	9.584	ug/l	97
43) Isopropyl Acetate	6.348	43	153246	25.066	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
Data File : VX029929.D
Acq On : 05 Jul 2022 18:36
Operator : JC/MD
Sample : PB145991ZHE#15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB145991ZHE#15

Quant Time: Jul 06 06:58:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

