

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
 Data File : VX029930.D
 Acq On : 05 Jul 2022 18:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 06 06:59:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.488	0.501	-2.7	94	0.00
3 P	Chloromethane	0.562	0.520	7.5	86	0.00
4 C	Vinyl Chloride	0.559	0.572	-2.3#	94	0.00
5 T	Bromomethane	0.308	0.224	27.3#	73	0.00
6 T	Chloroethane	0.327	0.282	13.8	86	0.00
7 T	Trichlorofluoromethane	0.717	0.732	-2.1	96	0.00
8 T	Diethyl Ether	0.322	0.312	3.1	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.501	0.484	3.4	95	0.00
10 T	Methyl Iodide	0.608	0.505	16.9	71	0.00
11 T	Tert butyl alcohol	0.166	0.162	2.4	94	0.01
12 CM	1,1-Dichloroethene	0.515	0.508	1.4#	96	0.00
13 T	Acrolein	0.083	0.048	42.2#	58	0.00
14 T	Allyl chloride	0.950	0.916	3.6	92	0.00
15 T	Acrylonitrile	0.386	0.384	0.5	94	0.00
16 T	Acetone	0.310	0.313	-1.0	98	0.00
17 T	Carbon Disulfide	1.280	1.182	7.7	88	0.00
18 T	Methyl Acetate	0.955	0.965	-1.0	98	0.00
19 T	Methyl tert-butyl Ether	1.738	1.709	1.7	92	0.00
20 T	Methylene Chloride	0.654	0.600	8.3	94	0.00
21 T	trans-1,2-Dichloroethene	0.565	0.542	4.1	94	0.00
22 T	Diisopropyl ether	1.920	1.876	2.3	92	0.00
23 T	Vinyl Acetate	1.661	1.648	0.8	91	0.00
24 P	1,1-Dichloroethane	1.013	1.014	-0.1	95	0.00
25 T	2-Butanone	0.537	0.540	-0.6	95	0.00
26 T	2,2-Dichloropropane	0.665	0.508	23.6	72	0.00
27 T	cis-1,2-Dichloroethene	0.667	0.660	1.0	95	0.00
28 T	Bromochloromethane	0.406	0.384	5.4	91	0.00
29 T	Tetrahydrofuran	0.354	0.356	-0.6	94	0.00
30 C	Chloroform	0.993	0.988	0.5#	94	0.00
31 T	Cyclohexane	0.882	0.869	1.5	94	0.00
32 T	1,1,1-Trichloroethane	0.791	0.799	-1.0	97	0.00
33 S	1,2-Dichloroethane-d4	0.633	0.625	1.3	93	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.312	0.309	1.0	97	0.00
36 T	1,1-Dichloropropene	0.416	0.390	6.2	93	0.00
37 T	Ethyl Acetate	0.568	0.546	3.9	92	0.00
38 T	Carbon Tetrachloride	0.360	0.363	-0.8	98	0.00
39 T	Methylcyclohexane	0.502	0.478	4.8	92	0.00
40 TM	Benzene	1.344	1.314	2.2	96	0.00
41 T	Methacrylonitrile	0.302	0.294	2.6	93	0.00
42 TM	1,2-Dichloroethane	0.412	0.403	2.2	93	0.00
43 T	Isopropyl Acetate	0.856	0.834	2.6	93	0.00
44 TM	Trichloroethene	0.362	0.341	5.8	99	0.00
45 C	1,2-Dichloropropane	0.355	0.342	3.7#	94	0.00
46 T	Dibromomethane	0.232	0.231	0.4	95	0.00
47 T	Bromodichloromethane	0.416	0.411	1.2	94	0.00
48 T	Methyl methacrylate	0.406	0.393	3.2	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
 Data File : VX029930.D
 Acq On : 05 Jul 2022 18:59
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 06 06:59:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 05:54:51 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.010	0.010	0.0	97	0.01
50 S	Toluene-d8	1.172	1.200	-2.4	99	0.00
51 T	4-Methyl-2-Pentanone	0.564	0.577	-2.3	97	0.00
52 CM	Toluene	0.811	0.802	1.1#	95	0.00
53 T	t-1,3-Dichloropropene	0.444	0.437	1.6	91	0.00
54 T	cis-1,3-Dichloropropene	0.501	0.488	2.6	91	0.00
55 T	1,1,2-Trichloroethane	0.353	0.348	1.4	95	0.00
56 T	Ethyl methacrylate	0.557	0.555	0.4	93	0.00
57 T	1,3-Dichloropropane	0.568	0.555	2.3	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.266	0.247	7.1	89	0.00
59 T	2-Hexanone	0.438	0.447	-2.1	96	0.00
60 T	Dibromochloromethane	0.312	0.319	-2.2	96	0.00
61 T	1,2-Dibromoethane	0.361	0.356	1.4	96	0.00
62 S	4-Bromofluorobenzene	0.414	0.444	-7.2	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.302	0.325	-7.6	110	0.00
65 PM	Chlorobenzene	1.004	0.982	2.2	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.335	0.341	-1.8	95	0.00
67 C	Ethyl Benzene	1.746	1.714	1.8#	96	0.00
68 T	m/p-Xylenes	0.670	0.667	0.4	97	0.00
69 T	o-Xylene	0.678	0.665	1.9	95	0.00
70 T	Styrene	1.121	1.116	0.4	96	0.00
71 P	Bromoform	0.239	0.258	-7.9	96	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.618	3.336	7.8	96	0.00
74 T	N-amyl acetate	1.791	1.701	5.0	92	0.00
75 P	1,1,2,2-Tetrachloroethane	1.424	1.278	10.3	93	0.00
76 T	1,2,3-Trichloropropane	1.197	1.141	4.7	96	0.00
77 T	Bromobenzene	0.870	0.807	7.2	96	0.00
78 T	n-propylbenzene	4.267	4.000	6.3	96	0.00
79 T	2-Chlorotoluene	2.537	2.403	5.3	98	0.00
80 T	1,3,5-Trimethylbenzene	3.003	2.788	7.2	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.377	0.356	5.6	87	0.00
82 T	4-Chlorotoluene	2.890	2.747	4.9	100	0.00
83 T	tert-Butylbenzene	2.977	2.760	7.3	95	0.00
84 T	1,2,4-Trimethylbenzene	3.009	2.840	5.6	97	0.00
85 T	sec-Butylbenzene	3.751	3.588	4.3	98	0.00
86 T	p-Isopropyltoluene	2.994	2.913	2.7	98	0.00
87 T	1,3-Dichlorobenzene	1.671	1.593	4.7	101	0.00
88 T	1,4-Dichlorobenzene	1.705	1.580	7.3	99	0.00
89 T	n-Butylbenzene	2.814	2.664	5.3	97	0.00
90 T	Hexachloroethane	0.488	0.495	-1.4	103	0.00
91 T	1,2-Dichlorobenzene	1.641	1.553	5.4	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.298	0.302	-1.3	103	0.00
93 T	1,2,4-Trichlorobenzene	0.995	0.896	9.9	94	0.00
94 T	Hexachlorobutadiene	0.368	0.302	17.9	89	0.00
95 T	Naphthalene	4.038	3.835	5.0	95	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
Data File : VX029930.D
Acq On : 05 Jul 2022 18:59
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jul 06 06:59:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 05:54:51 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.018	0.924	9.2	93	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6