



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070522\
Data File : VX029909.D
Acq On : 05 Jul 2022 10:00
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062822W.M
Title : SW846 8260
Last Update : Wed Jun 29 05:54:51 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.9	11448	PASS
75	95	30	60	45.4	30827	PASS
95	95	100	100	100.0	67885	PASS
96	95	5	9	6.6	4512	PASS
173	174	0.00	2	0.9	466	PASS
174	95	50	100	74.6	50627	PASS
175	174	5	9	8.0	4031	PASS
176	174	95	101	96.7	48955	PASS
177	176	5	9	6.2	3057	PASS