

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070523\
 Data File : VX036401.D
 Acq On : 05 Jul 2023 14:08
 Operator : JC/MD
 Sample : PB153983ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153983ZHE#02

Quant Time: Jul 06 04:50:57 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	167305	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	287103	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	255003	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	118268	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	108180	45.109	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	90.220%		
35) Dibromofluoromethane	5.385	113	90400	47.753	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	95.500%		
50) Toluene-d8	8.652	98	339761	48.052	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	96.100%		
62) 4-Bromofluorobenzene	11.079	95	121105	48.626	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	97.260%		
Target Compounds						Qvalue
16) Acetone	2.385	43	101909	99.418	ug/l	94
20) Methylene Chloride	2.788	84	7222	3.272	ug/l	86
25) 2-Butanone	4.556	43	45022	30.569	ug/l	98
37) Ethyl Acetate	4.720	43	31898	11.091	ug/l	97
43) Isopropyl Acetate	6.342	43	179279	39.074	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	307404	103.994	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070523\
Data File : VX036401.D
Acq On : 05 Jul 2023 14:08
Operator : JC/MD
Sample : PB153983ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153983ZHE#02

Quant Time: Jul 06 04:50:57 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

