

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070523\
 Data File : VX036413.D
 Acq On : 05 Jul 2023 18:53
 Operator : JC/MD
 Sample : PB153983ZHE#14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153983ZHE#14

Quant Time: Jul 06 04:56:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	183464	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	309611	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	274592	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	126983	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	117101	44.528	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	=	89.060%	
35) Dibromofluoromethane	5.385	113	96173	47.109	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	94.220%	
50) Toluene-d8	8.647	98	368869	48.377	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	=	96.760%	
62) 4-Bromofluorobenzene	11.079	95	133660	49.766	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	99.540%	
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	228816	203.561	ug/l	93
18) Methyl Acetate	2.709	43	5212	1.128	ug/l #	91
20) Methylene Chloride	2.794	84	7820	3.231	ug/l	90
25) 2-Butanone	4.574	43	125608	77.773	ug/l	98
37) Ethyl Acetate	4.721	43	55329	17.839	ug/l	98
43) Isopropyl Acetate	6.348	43	205642	41.562	ug/l	98
51) 4-Methyl-2-Pentanone	8.574	43	548642	172.111	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070523\
Data File : VX036413.D
Acq On : 05 Jul 2023 18:53
Operator : JC/MD
Sample : PB153983ZHE#14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153983ZHE#14

Quant Time: Jul 06 04:56:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

