

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070523\
 Data File : VX036414.D
 Acq On : 05 Jul 2023 19:17
 Operator : JC/MD
 Sample : PB153931TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB153931TB

Quant Time: Jul 06 04:56:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	181398	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	305850	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	280106	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	128432	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	118332	45.509	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	91.020%		
35) Dibromofluoromethane	5.385	113	98017	48.603	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.200%		
50) Toluene-d8	8.647	98	374284	49.690	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.380%		
62) 4-Bromofluorobenzene	11.079	95	133508	50.321	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	100.640%		
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	85718	77.126	ug/l	94
20) Methylene Chloride	2.794	84	4951	2.069	ug/l #	92
25) 2-Butanone	4.581	43	40705	25.491	ug/l	99
37) Ethyl Acetate	4.727	43	24224	7.906	ug/l	98
43) Isopropyl Acetate	6.342	43	169112	34.599	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070523\
Data File : VX036414.D
Acq On : 05 Jul 2023 19:17
Operator : JC/MD
Sample : PB153931TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB153931TB

Quant Time: Jul 06 04:56:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

