

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070523\
 Data File : VX036415.D
 Acq On : 05 Jul 2023 19:40
 Operator : JC/MD
 Sample : 03492-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-TP1

Quant Time: Jul 06 04:57:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 23 15:03:28 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	180173	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	306765	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	274928	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	124966	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	116317	45.038	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.080%
35) Dibromofluoromethane	5.391	113	95536	47.231	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.460%
50) Toluene-d8	8.647	98	364767	48.282	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.560%
62) 4-Bromofluorobenzene	11.079	95	130644	49.094	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.180%
Target Compounds						
					Qvalue	
16) Acetone	2.404	43	76162	68.994	ug/l	95
20) Methylene Chloride	2.788	84	6488	2.730	ug/l	92
25) 2-Butanone	4.580	43	40176	25.330	ug/l	93
37) Ethyl Acetate	4.727	43	21760	7.081	ug/l	98
43) Isopropyl Acetate	6.342	43	155800	31.781	ug/l	98
51) 4-Methyl-2-Pentanone	8.573	43	240769	76.231	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070523\
Data File : VX036415.D
Acq On : 05 Jul 2023 19:40
Operator : JC/MD
Sample : 03492-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-TP1

Quant Time: Jul 06 04:57:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X062323W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 23 15:03:28 2023
Response via : Initial Calibration

