

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070524\
 Data File : VX042205.D
 Acq On : 05 Jul 2024 11:48
 Operator : JC/MD
 Sample : P3167-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VPB196-HYD-20240703

Quant Time: Jul 05 23:20:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	145732	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	222468	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	105732	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69840	41.560	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	83.120%
35) Dibromofluoromethane	5.385	113	67901	43.015	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.040%
50) Toluene-d8	8.647	98	263376	47.494	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.980%
62) 4-Bromofluorobenzene	11.079	95	97487	47.690	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.380%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	2406	3.261	ug/l	# 84
20) Methylene Chloride	2.782	84	669	0.414	ug/l	91
60) Dibromochloromethane	9.525	129	1683	1.033	ug/l	99
71) Bromoform	10.799	173	938	1.496	ug/l	# 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070524\
Data File : VX042205.D
Acq On : 05 Jul 2024 11:48
Operator : JC/MD
Sample : P3167-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VPB196-HYD-20240703

Quant Time: Jul 05 23:20:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

