

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070524\
 Data File : VX042206.D
 Acq On : 05 Jul 2024 12:11
 Operator : JC/MD
 Sample : P3167-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-196-GW-865-867

Quant Time: Jul 05 23:20:28 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	141307	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	217826	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	216763	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101266	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68447	42.006	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	84.020%
35) Dibromofluoromethane	5.385	113	67309	43.549	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.100%
50) Toluene-d8	8.647	98	256939	47.321	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.640%
62) 4-Bromofluorobenzene	11.079	95	95067	47.497	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	95.000%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	4804	6.714	ug/l	# 84
17) Carbon Disulfide	2.501	76	1327	0.534	ug/l	99
20) Methylene Chloride	2.788	84	1234	0.788	ug/l	# 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070524\
Data File : VX042206.D
Acq On : 05 Jul 2024 12:11
Operator : JC/MD
Sample : P3167-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-196-GW-865-867

Quant Time: Jul 05 23:20:28 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

