

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070524\
 Data File : VX042209.D
 Acq On : 05 Jul 2024 13:22
 Operator : JC/MD
 Sample : P3167-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-196-TB-20240627

Quant Time: Jul 05 23:21:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	133147	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	205492	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	203872	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	93342	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66623	43.393	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.780%
35) Dibromofluoromethane	5.385	113	64545	44.267	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.540%
50) Toluene-d8	8.647	98	240720	46.995	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.980%
62) 4-Bromofluorobenzene	11.079	95	88877	47.070	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.140%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1503	2.229	ug/l	93
20) Methylene Chloride	2.782	84	1911	1.295	ug/l #	77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070524\
Data File : VX042209.D
Acq On : 05 Jul 2024 13:22
Operator : JC/MD
Sample : P3167-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-196-TB-20240627

Quant Time: Jul 05 23:21:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

