

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070524\
 Data File : VX042210.D
 Acq On : 05 Jul 2024 13:45
 Operator : JC/MD
 Sample : P3167-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-196-EB-20240701

Quant Time: Jul 05 23:21:59 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 21 23:51:58 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	131447	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	202341	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	197328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90482	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64481	42.541	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.080%
35) Dibromofluoromethane	5.391	113	62839	43.768	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.540%
50) Toluene-d8	8.647	98	233126	46.221	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	92.440%
62) 4-Bromofluorobenzene	11.079	95	85388	45.926	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.860%
Target Compounds						
16) Acetone	2.386	43	2335	3.508	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX070524\
Data File : VX042210.D
Acq On : 05 Jul 2024 13:45
Operator : JC/MD
Sample : P3167-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-196-EB-20240701

Quant Time: Jul 05 23:21:59 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 21 23:51:58 2024
Response via : Initial Calibration

