

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070618\
 Data File : VX003215.D
 Acq On : 06 Jul 2018 14:45
 Operator : JC/MD
 Sample : J3894-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 001M-WILLETS-PT-BLVD(JULY)

Quant Time: Jul 09 09:33:40 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X061818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jun 18 15:18:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	50231	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	265914	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	250029	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	109855	28.64	ug/l	0.00
Spiked Amount 30.000	Range 50 - 169		Recovery =	95.47%		
60) 4-Bromofluorobenzene	11.14	95	121990	29.54	ug/l	0.00
Spiked Amount 30.000	Range 56 - 143		Recovery =	98.47%		
63) Toluene-d8	8.72	98	312027	28.57	ug/l	0.00
Spiked Amount 30.000	Range 66 - 137		Recovery =	95.23%		

Target Compounds

					Ovalue
15) Acetone	2.45	58	957454	2173.283 ug/l	94
22) cis-1,2-Dichloroethene	4.60	96	9355	3.125 ug/l	89
30) 2-Butanone	4.71	43	90443	44.015 ug/l	97
58) 4-Methyl-2-Pentanone	8.65	43	18683	4.401 ug/l	99
62) Toluene	8.79	91	5744048	476.609 ug/l	99
79) tert-butylbenzene	12.07	119	527147	39.723 ug/l	98
82) p-Isopropyltoluene	12.07	119	527147	39.723 ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070618\
Data File : VX003215.D
Acq On : 06 Jul 2018 14:45
Operator : JC/MD
Sample : J3894-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
001M-WILLETS-PT-BLVD(JULY)

Quant Time: Jul 09 09:33:40 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X061818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jun 18 15:18:02 2018
Response via : Initial Calibration

