

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070618\
 Data File : VX003218.D
 Acq On : 06 Jul 2018 16:20
 Operator : JC/MD
 Sample : J3894-02DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 002M-35TH-AVE(JULY)DL

Quant Time: Jul 09 09:40:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA X\METHOD\624X061818W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Mon Jun 18 15:18:02 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.03	128	50756	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	6.87	114	265891	30.00	ug/l	0.00
57) Chlorobenzene-d5	10.12	117	247491	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	6.07	65	111677	28.81	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	96.03%
60) 4-Bromofluorobenzene	11.14	95	113050	27.66	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	92.20%
63) Toluene-d8	8.72	98	309639	28.64	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	95.47%

Target Compounds

					Ovalue
15) Acetone	2.45	58	70213	157.725	ug/l 94
62) Toluene	8.79	91	486941	40.818	ug/l 98
79) tert-butylbenzene	12.07	119	33649	2.562	ug/l 99
82) p-Isopropyltoluene	12.07	119	33649	2.562	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070618\
Data File : VX003218.D
Acq On : 06 Jul 2018 16:20
Operator : JC/MD
Sample : J3894-02DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
002M-35TH-AVE(JULY)DL

Quant Time: Jul 09 09:40:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA X\METHOD\624X061818W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Mon Jun 18 15:18:02 2018
Response via : Initial Calibration

