

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070618\
 Data File : VX003221.D
 Acq On : 06 Jul 2018 19:20
 Operator : JC/MD
 Sample : VX0706WBL02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0706WBL02

Quant Time: Jul 07 00:12:05 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	284203	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	395143	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	366829	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	204495	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	179570	45.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.42%	
35) Dibromofluoromethane	5.50	113	153745	49.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.26%	
50) Toluene-d8	8.72	98	546715	47.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.68%	
62) 4-Bromofluorobenzene	11.14	95	191060	43.35	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.70%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070618\
Data File : VX003221.D
Acq On : 06 Jul 2018 19:20
Operator : JC/MD
Sample : VX0706WBL02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX0706WBL02

Quant Time: Jul 07 00:12:05 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

