

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070618\  
 Data File : VX003226.D  
 Acq On : 06 Jul 2018 21:33  
 Operator : JC/MD  
 Sample : J3754-08DL 10X  
 Misc : 5.0mL/MSVOA X/WATER  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 S10-0145DL

Quant Time: Jul 09 09:29:05 2018  
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA\_X\METHOD\82X062518W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jun 25 14:38:05 2018  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units  | Dev(Min) |
|-----------------------------|-------|------|----------|--------|--------|----------|
| 1) Pentafluorobenzene       | 5.67  | 168  | 274212   | 50.00  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.87  | 114  | 382946   | 50.00  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5        | 10.12 | 117  | 361648   | 50.00  | ug/l   | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.08 | 152  | 198663   | 50.00  | ug/l   | 0.00     |
| System Monitoring Compounds |       |      |          |        |        |          |
| 33) 1,2-Dichloroethane-d4   | 6.07  | 65   | 175384   | 46.27  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 92.54% |          |
| 35) Dibromofluoromethane    | 5.51  | 113  | 146381   | 48.27  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 96.54% |          |
| 50) Toluene-d8              | 8.72  | 98   | 529424   | 47.81  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 95.62% |          |
| 62) 4-Bromofluorobenzene    | 11.14 | 95   | 189563   | 44.39  | ug/l   | 0.00     |
| Spiked Amount 50.000        |       |      | Recovery | =      | 88.78% |          |
| Target Compounds            |       |      |          |        |        |          |
| 11) Tert butyl alcohol      | 3.08  | 59   | 12590    | 16.675 | ug/l # | 94       |
| 40) Benzene                 | 6.15  | 78   | 830372   | 68.053 | ug/l   | 100      |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070618\  
Data File : VX003226.D  
Acq On : 06 Jul 2018 21:33  
Operator : JC/MD  
Sample : J3754-08DL 10X  
Misc : 5.0mL/MSVOA X/WATER  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
S10-0145DL

Quant Time: Jul 09 09:29:05 2018  
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA\_X\METHOD\82X062518W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jun 25 14:38:05 2018  
Response via : Initial Calibration

