

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070618\
 Data File : VX003230.D
 Acq On : 06 Jul 2018 23:20
 Operator : JC/MD
 Sample : J3846-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WP022RW350-180702

Quant Time: Jul 09 09:21:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 25 14:38:05 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	279592	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.87	114	390118	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	364677	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	200275	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	177980	46.05	ug/l	0.00
Spiked Amount 50.000			Recovery =	92.10%		
35) Dibromofluoromethane	5.51	113	150961	48.86	ug/l	0.00
Spiked Amount 50.000			Recovery =	97.72%		
50) Toluene-d8	8.72	98	540396	47.90	ug/l	0.00
Spiked Amount 50.000			Recovery =	95.80%		
62) 4-Bromofluorobenzene	11.14	95	197728	45.45	ug/l	0.00
Spiked Amount 50.000			Recovery =	90.90%		

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.40	62	11214	3.547	ug/l #	84
11) Tert butyl alcohol	3.07	59	22845	29.675	ug/l	96
24) 1,1-Dichloroethane	3.70	63	18856	3.339	ug/l	99
27) cis-1,2-Dichloroethene	4.60	96	16108	4.784	ug/l	94
40) Benzene	6.15	78	24182	1.945	ug/l #	89
42) 1,2-Dichloroethane	6.20	62	29915	5.958	ug/l	95
52) Toluene	8.79	92	3292	0.413	ug/l	98
65) Chlorobenzene	10.14	112	1043782	112.686	ug/l	99
87) 1,3-Dichlorobenzene	12.03	146	21104	3.012	ug/l	98
88) 1,4-Dichlorobenzene	12.10	146	61156	8.480	ug/l	97
91) 1,2-Dichlorobenzene	12.40	146	82253	11.630	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070618\
Data File : VX003230.D
Acq On : 06 Jul 2018 23:20
Operator : JC/MD
Sample : J3846-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WP022RW350-180702

Quant Time: Jul 09 09:21:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062518W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 25 14:38:05 2018
Response via : Initial Calibration

