

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070620\
 Data File : VX017163.D
 Acq On : 06 Jul 2020 13:24
 Operator : JC/SP
 Sample : VX0706MBL01
 Misc : 5.0g/10mL/100uL/5.0mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0706MBL01

Quant Time: Jul 06 15:44:40 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	269445	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	435621	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	450914	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	246912	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	167767	50.51	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.02%	
35) Dibromofluoromethane	5.47	113	140340	50.93	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.86%	
50) Toluene-d8	8.70	98	531394	52.15	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.30%	
62) 4-Bromofluorobenzene	11.12	95	222220	55.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	111.58%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070620\
Data File : VX017163.D
Acq On : 06 Jul 2020 13:24
Operator : JC/SP
Sample : VX0706MBL01
Misc : 5.0µ/10mL/100uL/5.0mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0706MBL01

Quant Time: Jul 06 15:44:40 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

