

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070620\
 Data File : VX017171.D
 Acq On : 06 Jul 2020 16:26
 Operator : JC/SP
 Sample : L3159-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 63BR-20200629

Quant Time: Jul 07 07:27:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	277715	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	460479	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	477950	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	266365	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	170092	49.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.36%	
35) Dibromofluoromethane	5.47	113	149656	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
50) Toluene-d8	8.70	98	557081	51.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.44%	
62) 4-Bromofluorobenzene	11.12	95	243924	57.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.86%	
Target Compounds						
16) Acetone	2.42	43	8232	6.427	ug/l	95
27) cis-1,2-Dichloroethene	4.57	96	4987	1.443	ug/l	88
79) 2-Chlorotoluene	11.40	91	4218	0.353	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070620\
Data File : VX017171.D
Acq On : 06 Jul 2020 16:26
Operator : JC/SP
Sample : L3159-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
63BR-20200629

Quant Time: Jul 07 07:27:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

