

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX070620\
 Data File : VX017172.D
 Acq On : 06 Jul 2020 16:48
 Operator : JC/SP
 Sample : L3159-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40S-20200629

Quant Time: Jul 07 07:29:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jun 29 17:10:38 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	282504	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	467607	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	486065	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	270309	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	176792	50.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.52%	
35) Dibromofluoromethane	5.47	113	154108	52.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.20%	
50) Toluene-d8	8.70	98	570540	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
62) 4-Bromofluorobenzene	11.12	95	241654	56.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.04%	
Target Compounds						
16) Acetone	2.42	43	7299	5.602	ug/l	94
21) trans-1,2-Dichloroethene	3.15	96	3180	1.025	ug/l	85
27) cis-1,2-Dichloroethene	4.57	96	40376	11.489	ug/l	89
44) Trichloroethene	7.19	130	46813	12.834	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX070620\
Data File : VX017172.D
Acq On : 06 Jul 2020 16:48
Operator : JC/SP
Sample : L3159-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40S-20200629

Quant Time: Jul 07 07:29:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X062920W.M
Quant Title : SW846 8260
QLast Update : Mon Jun 29 17:10:38 2020
Response via : Initial Calibration

